

REPLIES FOR PREBID QUERIES FOR T NO 129 (SURGICAL CONSUMABLES)

| Firm Name | Item Code | Item Name | Firm Remarks | Recommended Generic Item Name & Specification | As per Tender | Final Remarks |
|-------------------------|-----------|--|--|---|---|----------------|
| India Medtronic Pvt Ltd | 2021090 | 60mm cartridges for laproscopic battery operated endocutter with Built in knife & articulation with gripping surface technology blue,gold& green reloads | This Specification is a lock-in specification & it's favoring only single company/ manufacturer. With Generic specification, the Corporation will get more no. of bidders. | <p>Proposed Item Name-</p> <p>60mm cartridges for laparoscopic endocutter with Built in knife & articulation with white, blue & green reloads</p> <p>Proposed Specification-</p> <p>Endoscopic linear cutter 60mm cartridge for consistent staple formation having 6 rows of staple line with 3 on each side of cut line.</p> | Endoscopic linear cutter 60mm cartridge with surface having pocket extensions to prevent tissue milking, staples with asymmetric bent tip for consistent staple formation with an option of closed staple height of 1.5, 1.8, 2mm & 6 rows of staple line with 3 on each side of cut line. | Not Considered |
| | 2021020 | Absorbable Hearnia Fixation Device for Laparoscopic with 25 straps (Tachers) | This Specification is a lock-in specification & it's favoring only single company/ manufacturer. With Generic specification, the Corporation will get more no. of bidders. | <p>Proposed Item name : Absorbable Hernia Mesh Fixation Device for Laparoscopic procedures with 25-30 straps/tacks</p> <p>Proposed Specification: Absorbable Hernia mesh fixation device preloaded with 25-30 absorbable straps/tacks having smooth mechanism of triggering, USFDA approved.</p> | Two-point absorbable Fixation device for Lap hernia repair. Trigger squeeze mechanical device with straight cannula, pre-loaded with 25 absorbable straps which are made of a blend of polydiaxanone dyed with D&C Violet No. 2 and L(-)-lactide/Glycolide copolymer. Inserted length of strap should be 6.7 mm USFDA approved. | Not Considered |
| | 2021111 | Absorbable, antibacterial Poliglecaprone 25, Sterilized Monofilament surgical suture, size 3-0 (Box of 12 Foils) | This Specification is a lock-in specification & it's favoring only single company/ manufacturer. With Generic specification, the Corporation will get more no. of bidders. | <p>Proposed Item name:</p> <p>Absorbable Poliglecaprone-25 / Polyglytone-6211, Sterilized Monofilament surgical suture, size 3-0</p> <p>Proposed Specification :</p> | Triclosan Antibacterial coated Monofilament Poliglecaprone 25 suture, length 70 cm ,size 3- 0 with 3/8 reverse cutting, , 26mm needle | Not Considered |

| Firm Name | Item Code | Item Name | Firm Remarks | Recommended Generic Item Name & Specification | As per Tender | Final Remarks |
|-----------|-----------|---|--|---|---|----------------|
| | | | | Undyed Monofilament Poliglecaprone-25 / Polyglytone-6211, suture length 70-75 cm, size 3-0 with 3/8 reverse cutting, 26mm needle | | |
| | 2021099 | Antibacterial Absorbable Surgical suture Polyglactin 910 1/2 circle reverse cutting 36mm OS Needle violet brided coated - 90cm, size-1- 0.with Triclosan coating. (Box of 12 Foils) | This Specification is a lock-in specification & it's favoring only single company/ manufacturer. With Generic specification, the Corporation will get more no. of bidders. | <p>Proposed Item name : Coated Absorbable Surgical Suture Polyglactin 910 1/2 Circle reverse cutting 36-38mm OS Needle violet braided coated - 90cm, size 1-0</p> <p>Proposed Specification : Coated Absorbable Surgical suture Polyglactin 910, 1/2 circle reverse cutting 36-38mm OS Needle, violet braided- 90cm, size 1-0</p> | Antibacterial Absorbable Surgical suture Polyglactin 910 1/2 circle reverse cutting 36mm OS Needle violet brided coated - 90cm, size- 1-0.with Triclosan coating. | Not Considered |
| | 2021098 | Antibacterial Absorbable Surgical suture Polyglactin 910 1/2 circle reverse cutting 40mm OS Needle violet brided coated - 90cm, size- 1.with Triclosan coating.(Box of 12 Foils) | This Specification is a lock-in specification & it's favoring only single company/ manufacturer. With Generic specification, the Corporation will get more no. of bidders. | <p>Proposed Item Name :</p> <p>Coated Absorbable Surgical Suture Polyglactin 910 ,1/2 Circle reverse cutting 40mm OS Needle violet braided coated - 90cm, size 1</p> <p>Proposed Specification :</p> <p>Coated Absorbable Surgical suture Polyglactin 910, 1/2 circle reverse cutting 40mm OS Needle, violet braided - 90cm,</p> | Antibacterial Absorbable Surgical suture Polyglactin 910 1/2 circle reverse cutting 40mm OS Needle violet brided coated - 90cm, size- 1.with Triclosan coating. | Not Considered |

| Firm Name | Item Code | Item Name | Firm Remarks | Recommended Generic Item Name & Specification | As per Tender | Final Remarks |
|-----------|-----------|---|--|---|---|----------------|
| | | | | size 1 | | |
| | 2021088 | Fixed Envil Hemorrhoid Stapler kit | This Specification is a lock-in specification & it's favoring only single company/ manufacturer. With Generic specification, the Corporation will get more no. of bidders. | Proposed Item Name- Haemorrhoid Stapler Kit Proposed Specification- Haemorrhoid stapler set with 2rows of staple line | Fixed Envil Hemorrhoid stapler Kit with 2rows and closed staple height range from 0.75mm to 1.5mm | Not Considered |
| | 2021021 | Hernia mesh fixation device with absorbable straps for Open Hernia(Tachers) | This Specification is a lock-in specification & it's favoring only single company/ manufacturer. With Generic specification, the Corporation will get more no. of bidders. | Proposed Item name : Hernia mesh fixation device with absorbable straps/tacks for Open Hernia (Tackers) Proposed Specification : Absorbable Hernia mesh fixation device preloaded with 15-20 absorbable straps/tacks having smooth mechanism of triggering for Open Hernia repair surgery | 2 point fixation device for open hernia repairs with a curved cannula & strap- positioning tip having a forward-tilted handle & metric ruler. It is a trigger-squeeze mechanical device with a curved cannula, pre- loaded with 20 absorbable straps which are made of a blend of polydioxanone dyed with D&C Violet No. 2 and an L(-)- lactide/glycolide copolymer. Inserted length of strap should be 6.7 mm. | Not Considered |

| Firm Name | Item Code | Item Name | Firm Remarks | Recommended Generic Item Name & Specification | As per Tender | Final Remarks |
|-----------|-----------|---|--|--|--|---|
| | 2021089 | Laprosopic 60mm battery operated endocutter with Built in knife & articulation. | This Specification is a lock-in specification & it's favoring only single company/ manufacturer. With Generic specification, the Corporation will get more no. of bidders. | Proposed Item Name- Laparoscopic 60mm endocutter with Built in knife & articulation Proposed Specification- 60 mm Articulating Endo cutter with 6 rows (3 on either side of cut line) articulation up to 45 degrees on both sides capable of entry through a 12mm trocar | Light weight Battery operated 60 mm Articulating Endo cutter with a disposable battery pack, 360 degree rotation shaft, 6 row(3 on either side of cut line) staples offering 60mm staple line and 57 mm cut line, one handed natural articulation up to 45 degrees on both sides, precision machined closed channel anvil, capable of delivering system wide pre compression, wide proximal to distal jaw aperture (proximal 8 mm, distal 22 mm), 3 point gap control for alignment and calibration throughout the 60 mm staple line, manual override safety feature, interchangeable 6 row(3 on either side of cut line) cartridge options of 1.5,1.8,& 2.0mm closed staple height, capable of entry through a 12 mmtrocar sleeve, shaft length of 340mm. | Not Considered |
| | 2021059 | Laprosopic preloaded 5mm dia clip applier with medium large with 15 clips. | This Specification is a lock-in specification & it's favoring only single company/ manufacturer. With Generic specification, the Corporation will get more no. of bidders. | Proposed Item Name- Laparoscopic preloaded 5mm dia clip applier with medium large with 15-16 clips. Proposed Specification- Preloaded multiple clip applier 5 mm with 15-16 titanium clips of medium-large size for laparoscopic procedure capable of entry through 5mm trocar. | Endoscopic 360 degrees rotating multiple clip applier 5 mm with 15 pre loaded titanium clips of medium large size with shaft length of 33 mm, capable of entry through 5mm trocar, offering ratchet mechanism, low clip indicator, collapsable jaws and closed clip length of 8.8 mm, clips having lateral & transverse grooves for onvessel | Not Considered (Rachet mechanism helps in holding the clips) |

| Firm Name | Item Code | Item Name | Firm Remarks | Recommended Generic Item Name & Specification | As per Tender | Final Remarks |
|-----------|-----------|--|--|--|---|--|
| | | | | | security. | |
| | 2021087 | New Curved Circular Stapler with Contolled Tissue technology 29/33 mm | This Specification is a lock-in specification & it's favoring only single company/ manufacturer. With Generic specification, the Corporation will get more no. of bidders. | Proposed Item Name- Circular Stapler for end to end anastomosis sizes- 25/ 28/29/31/33 Proposed Specification- Circular stapler for end to end anastomosis sizes- 25/ 28/29/31/33 with Curved shaft anvil detachable head Non-slip griping surface for increased control during firing, larger tissue compression scale for visibility & ease of use. | Proximate Intra-Luminal Staplers with Curved shaft anvil detachable head (Available in diameter-29mm & 33mm) lumen size (14.4, 20.4, & 24.4mm respectively) with adjustable staple height (1.0-2.5 mm) for controlled tissue compression, longer staple leg 5.5mm, Non-slip griping surface for increased control during firing, larger tissue compression scale for visibility & ease of use. | Not Considered (Controlled tissue technology is good having range 1.0 -2.5mm) |
| | 2021023 | Open Flexible Composite Mesh Device is a sterile, partially absorbable skirted mesh device designed for the repair of open ventral hernias | This Specification is a lock-in specification & it's favoring only single company/ manufacturer. With Generic specification, the Corporation will get more no. of bidders. | Proposed Item Name : Open Flexible Composite Mesh Device is a sterile, partially absorbable skirted mesh device designed for the repair of open ventral hernias. Proposed Specification : Composed of a macroporous mesh knitted from polypropylene/polyester for hernia defect reinforcement and absorbable film for preventing visceral adhesions. Size 12X15 cm or 10X15 cm. | composed of a macroporous mesh, knitted from polypropylene and polydioxanone fibers, laminated to an absorbable poliglecaprone 25 film, size 12x15 | Not Considered |

| Firm Name | Item Code | Item Name | Firm Remarks | Recommended Generic Item Name & Specification | As per Tender | Final Remarks |
|-----------|-----------|-----------------------------------|--|---|--|---|
| | 2021082 | Open 3 Row Linear Stapler - 55 mm | This Specification is a lock-in specification & it's favoring only single company/ manufacturer. With Generic specification, the Corporation will get more no. of bidders. | Proposed Item Name- Open 3/2 Rows Linear Stapler - 55 mm/60mm Proposed Specification- Linear cutter 55mm/60mm with 3/2 Rows staple line, two-sided firing for easy handling for consistent staple formation | Linear cutter 55mm with intermediate locking position for easy tissue manipulation, two sided firing for easy handling, proximal pre load support for consistent staple formation, with selectable closed staple height of 1.5mm,1.8mm and 2mm | Not Considered (3 rows technologies is good) |
| | 2021084 | Open 3 Row Linear Stapler - 75mm | This Specification is a lock-in specification & it's favoring only single company/ manufacturer. With Generic specification, the Corporation will get more no. of bidders. | Proposed Item Name- Open 3/2 Rows Linear Stapler - 75 mm/80mm Proposed Specification- Linear cutter 75mm/80mm with 3/2 Rows staple line, two-sided firing for easy handling for consistent staple formation | Linear cutter 75mm with intermediate locking position for easy tissue manipulation, two sided firing for easy handling, proximal pre load support for consistent staple formation, with selectable closed staple height of 1.5mm,1.8mm and 2mm. | Not Considered (3 rows technologies is good and upgraded) |
| | 2021085 | Open Curved Cutter- 40mm | This Specification is a lock-in specification & it's favoring only single company/ manufacturer. With Generic specification, the Corporation will get more no. of bidders. | Proposed Item Name- Open Curved/L type Cutter- 40mm/45mm/60mm. Proposed Specification- Gastro-intestinal stapler with curved head/L shape designed for low pelvic access with enhanced visibility, delivering 40mm/45mm/60mm single cut line in between the staple line, tissue retaining pin to prevent tissue slippage, intermediate locking position for precise placement. | Gastro-intestinal stapler with curved head designed for low pelvic access with enhanced visibility, delivering four curvilinear rows of staples in the width of 30mm space with a 40 mm single cut line in between the staple line, tissue retaining pin to prevent tissue slippage,intermediate locking position for precise placement. | Not considered (curved 40mm is useful in pelvic surgeries) |

| Firm Name | Item Code | Item Name | Firm Remarks | Recommended Generic Item Name & Specification | As per Tender | Final Remarks |
|-----------|-----------|---|--|---|---|--|
| | 5574 | Optical view trocar (Laparoscopy) 11mm | This Specification is a lock-in specification & it's favoring only single company/ manufacturer. With Generic specification, the Corporation will get more no. of bidders. | Proposed Item Name- Optical view trocar (Laparoscopy) 11mm Proposed Specification- Optical view trocar 11 mm Bladeless | Optical view trocar 11 mm Bladeless | Both are same |
| | 2021106 | Optical view trocar (Laparoscopy) 12mm | This Specification is a lock-in specification & it's favoring only single company/ manufacturer. With Generic specification, the Corporation will get more no. of bidders. | Proposed Item Name- Optical view trocar (Laparoscopy) 12mm Proposed Specification- Blade less 12 mm Trocar with Optical view technology. | Blade less 12 mm Trocar with Optical view technology, should have wipe, wick and absorb features. | Not Considered (Wick and wipe technology is good) |
| | 5575 | Optical view trocar (Laparoscopy) 5mm | This Specification is a lock-in specification & it's favoring only single company/ manufacturer. With Generic specification, the Corporation will get more no. of bidders. | Proposed Item Name- Optical view trocar (Laparoscopy) 5mm Proposed Specification- Blade less 5 mm Trocar with Optical view technology. | Blade less 5 mm Trocar with Optical view technology. | Both are same |
| | 2021109 | PDS - Antibacterial, Absorbable Polydioxanone Sterilized Symmetric, Knotless Tissue Controlling Device, Size -1 (Box of 12 Foils) | This Specification is a lock-in specification & it's favoring only single company/ manufacturer. With Generic specification, the Corporation will get more no. of bidders. | Proposed Item Name : Sterilized Unidirectional Knotless Tissue Controlling Device, Size 1 Proposed Specification : Knotless Unidirectional Tissue controlling device, size 1, 45-60 cm, 1/2 circle, 37-40mm, Taper Point | Triclosan antibacterial coated Absorbable Polydioxanone Unidirectional with Fixation Tab, 40mm, 60cm1/2 Circle, Taper Point, CT | Not considered |

| Firm Name | Item Code | Item Name | Firm Remarks | Recommended Generic Item Name & Specification | As per Tender | Final Remarks |
|-----------|-----------|--|--|---|--|----------------|
| | 2021097 | PDS - Polydioxanone with Triclosan Antibacterial coated Violet Monofilament 90 cm size 1 (Box of 24 Foils) | This Specification is a lock-in specification & it's favoring only single company/ manufacturer. With Generic specification, the Corporation will get more no. of bidders. | Proposed Item Name : Polydioxanone/Polyglyconate coloured Monofilament suture - 90cm, size 1 Proposed Specification : Polydioxanone/Polyglyconate coloured Monofilament suture, 90cm, size 1 with 1/2 Circle, Taper Point, 40mm needle, USFDA approved. | Polydioxanone with Triclosan Antibacterial coated Violet Monofilament 90 cm size 1 with 1/2 Circle Taper Point CT-1, 40mm needle USFDA approved. | Not considered |
| | 2021096 | PDS - Polydioxanone with Triclosan Antibacterial coated Violet Monofilament 90 cm size 0 (Box of 36 Foils) | This Specification is a lock-in specification & it's favoring only single company/ manufacturer. With Generic specification, the Corporation will get more no. of bidders. | Proposed Item Name : Polydioxanone/Polyglyconate coloured Monofilament suture - 90cm, size 0 Proposed Specification : Polydioxanone/Polyglyconate coloured Monofilament suture, 90cm, size 0 with 1/2 Circle, Taper Point, 36-38mm needle, USFDA approved. | Polydioxanone with Triclosan Antibacterial coated Violet Monofilament 90 cm size 0 with 1/2 Circle Taper Point CT-1, 36.4mm needle USFDA approved. | Not considered |
| | 2021086 | Reloads For Open Curved Cutter-40mm Green. | This Specification is a lock-in specification & it's favoring only single company/ manufacturer. With Generic specification, the Corporation will get more no. of bidders. | Proposed Item Name- Reloads for Open Curved/L type Cutter- 40mm/45mm/60mm Proposed Specification- Curved/L shape designed green cartridge for low pelvic access having close staple height of 2.0 mm, tactile feedback on completion of firing sequence. | Curved green cartridge having close staple height of 2.0 mm, tactile feedback on completion of firing sequence, new anvil, knife with every cartridge. | Not considered |

| Firm Name | Item Code | Item Name | Firm Remarks | Recommended Generic Item Name & Specification | As per Tender | Final Remarks |
|-----------|-----------|--|--|---|--|----------------|
| | 5926 | Skin Stapler | This Specification is a lock-in specification & it's favoring only single company/ manufacturer. With Generic specification, the Corporation will get more no. of bidders. | Proposed Item Name- Skin Stapler Proposed Specification- Sterile, Disposable, 35W Stainless Steel | Sterile, Disposable, 35 W. Fine 31L. Stainless Steel | Not considered |
| | 5925 | Skin stapler removals | This Specification is a lock-in specification & it's favoring only single company/ manufacturer. With Generic specification, the Corporation will get more no. of bidders. | Proposed Item Name- Skin Stapler removals Proposed Specification- Skin stapler removals stainless steel | Skin stapler removals stainless steel | Not considered |
| | 2021022 | Tissue separating multilayer mesh for intra-abdominal open and laparoscopic ventral hernia repair. | This Specification is a lock-in specification & it's favoring only single company/ manufacturer. With Generic specification, the Corporation will get more no. of bidders. | Proposed Item Name : Tissue separating multilayer component mesh for intra-abdominal open and laparoscopic ventral hernia repair. Proposed Specification : Tissue separating multilayer component mesh for intra-abdominal open & laparoscopic ventral hernia repair. Soft polypropylene/polyester mesh clubbed with absorbable films on visceral side for preventing adhesions. Size 15 cm X 15 cm. USFDA approved. | Tissue separating multilayer mesh for intra-abdominal open and laparoscopic ventral hernia repair. Soft polypropylene mesh clubbed with polydioxanone layer and has an oxidized regenerated cellulose on visceral side.Sizes-15cmx15cm, USFDA approved | Not considered |
| | 2021072 | Enseal Open Probe with 20cms | Enseal Open probe with 20 cm is one company specific and brand name. Adding reconsideration will allow multiple companies to | Vessel Sealing device with Jaw length of 35-36 mm and 180 to 360 degree rotation Open probe with inbuilt blade for tissue division. | Enseal Open Probe with 20cms, with jaw length of 35mm, width 6mm and 360degree rotation | Not considered |

| Firm Name | Item Code | Item Name | Firm Remarks | Recommended Generic Item Name & Specification | As per Tender | Final Remarks |
|-----------|-----------|---|---|--|---|----------------|
| | | | participate in the tender. | | | |
| | 2021073 | Enseal G2 Tissue Sealer Curved jaw Lap35cms with articulation and 360 degree rotation | Enseal G2 Tissue Sealer curved Jaw 35cm is one company specific and brand name. Adding reconsideration will allow multiple companies to participate in the tender. | Vessel Sealing device Curved Jaw Lap with inbuilt blade for tissue division 35-37 cm and 359-360 degree rotation | Enseal G2 Tissue Sealer Curved jaw Lap35cms with articulation and 360 degree rotation | Not considered |
| | 2021094 | Harmonic Blue Handpiece For Open Surgeries | Blue handpiece is with one company only and color will not impact the end result of surgery. Adding reconsideration will allow multiple companies to quote. | Connecting transducer cum generator handpiece for ultrasonically activated cutting and coagulation device. | Connecting cable(Transducer/handpiece) Blue Handpiece for ultrasonically activatedcutting and coagulating device, | Not considered |
| | 2021093 | Harmonic open probe 23cm | Harmonic ACE +7 is a brand name and one company specific. Adding reconsideration will allow multiple companies to quote. | Ultrasonic cutting and coagulation shear 23-26 cm. | Harmonic ACE +7 wuth Advanced Hemostatis 23cm | Not considered |
| | 2021092 | Harmonic Probes for Open Cases - 17cm | Ultrasonic Open Shear Focus is brand name and Adaptive Tissue Technology is one company specific. Adding reconsideration with range in length will allow multiple bidders to participate. | Ultrasonic Open Shear for cutting and coagulation, 17- 26 cm long | Ultrasonic Open Shear Focus, 17cm long with Adaptive Tissue Technology | Not considered |

| Firm Name | Item Code | Item Name | Firm Remarks | Recommended Generic Item Name & Specification | As per Tender | Final Remarks |
|---------------------------|-----------|--|---|--|--|---|
| | 2021091 | Harmonic Probes for Open Cases - 9cm | Ultrasonic Open Shear Focus is brand name and Adaptive Tissue Technology is one company specific. Adding reconsideration with range in length will allow multiple bidders to participate. | Ultrasonic Open Shear for cutting and coagulation, 9- 13 cm long | Ultrasonic Open Shear Focus, 9cm long with Adaptive Tissue Technology | Not considered |
| | 2021074 | Harmonic HD1000i | Harmonic is a brand name and one company specific product. Adding reconsideration will allow multiple bidders to participate. | Ultrasonic Lap Probe 36-39 cm for cutting and coagulation | Harmonic Lap probe 36cm with integrated handpiece& advanced Hemostatis | Not considered |
| | 2021100 | Reusable Patient return electrode made of akton polymer for Monopolar | Reusable Patient return electrode made of akton polymer with current limiting nature technology is one company specific. Adding reconsideration will allow multiple bidders to participate. | Reusable Patient Return Electrode for all ESU. | Reusable Patient return electrode made of akton polymer with current limiting nature technology. 117cm(46")Length X 51cm(20")Width,1.25cm(1/2") thickness. | Not Considered (Avoid Pad side burn) |
| Johnson & Johnson Pvt Ltd | 2021090 | 60mm cartridges for laproscopic battery operated endocutter with Built in knife & articulation with gripping surface technology blue,gold& green reloads | Asked blue, gold & green in single item code, whereas 3 are different product codes – request you to clarify on how to quote the price bid | | Endoscopic linear cutter 60mm cartridge with surface having pocket extensions to prevent tissue milking, staples with asymmetric bent tip for consistent staple formation with an option of closed staple height of 1.5, 1.8, 2mm & 6 rows of staple line with 3 on each side of cut line. | Read as 2021090.01 Blue, 2021090.02 Gold, 2021090.03 Green |

| Firm Name | Item Code | Item Name | Firm Remarks | Recommended Generic Item Name & Specification | As per Tender | Final Remarks |
|-----------|-----------|---|---|---|---|---|
| | 2021087 | New Curved Circular Stapler with Contolled Tissue technology 29/33 mm | Asked 29 & 33mm in single item code, whereas both are different product codes – request you to clarify on how to quote the price bid. | | Proximate Intra-Luminal Staplers with Curved shaft anvil detachable head (Available in diameter-29mm & 33mm) lumen size (14.4, 20.4, & 24.4mm respectively) with adjustable staple height (1.0-2.5 mm) for controlled tissue compression, longer staple leg 5.5mm, Non-slip griping surface for increased control during firing, larger tissue compression scale for visibility & ease of use. | Read as 2021087.01 29mm 2021087.02 33mm |
| | 2021062 | Polyamide with cutting needle Size 1 | same item code has been awarded to us in Tender95 under item code 20211044 and RC is still valid, hence requesting you to remove this item code from current tender no 129. | | Size 1 | Deleted from the tender |

| Firm Name | Item Code | Item Name | Firm Remarks | Recommended Generic Item Name & Specification | As per Tender | Final Remarks |
|-----------|-----------|-----------|---|---|---------------|----------------|
| | | | <p>2. Market Standing: The items invited in this tender are very critical in treating patient, hence at least three years market standing should be available with participants so that the Corporation can be sure about the quality of the product. Therefore, requesting you to consider the Market standing issued by concerned licensing Authority of Drugs control Department in case of Drugs/Surgical Item or from CA in case of non -drug item for preceding 3 years. However, for “New Drugs” in accordance with Rule 122-E to the Drugs and Cosmetics Rules, 1945, relaxation for requirement of 3 years period should be given for “New Products”.</p> | | | Not considered |

| Firm Name | Item Code | Item Name | Firm Remarks | Recommended Generic Item Name & Specification | As per Tender | Final Remarks |
|---------------------------|-----------|--|--|---|---|----------------|
| | | | 3. Product Specification: The specifications for the items invited in this tender are very detailed and elaborated and product license will not have scope to accommodate entire specification, though they will mention the key specification. Hence, we request you to consider the other supporting regulatory documents like IFU to substantiate the specification. | | | Not considered |
| Healthium medtech limited | 10274 | 10/0 Proline suture double Arm (Box of 12) | Please specify the complete specification of the tendered product and remove the brand name | | 10/0 Proline suture double Arm | Not considered |
| | 2021099 | Antibacterial Absorbable Surgical suture Polyglactin 910 1/2 circle reverse cutting 36mm OS Needle violet bridged coated - 90cm, size-1- 0.with Triclosan coating. (Box of 12 Foils) | We would like request you to remove the "OS" from these products specification which is monopoly specification of other bidder. We request you to kindly remove the same from specification and accept the bid in general specification for Healthy competition. | | Antibacterial Absorbable Surgical suture Polyglactin 910 1/2 circle reverse cutting 36mm OS Needle violet bridged coated - 90cm, size- 1- 0.with Triclosan coating. | Not considered |

| Firm Name | Item Code | Item Name | Firm Remarks | Recommended Generic Item Name & Specification | As per Tender | Final Remarks |
|-----------|-----------|--|---|---|---|----------------|
| | 2021098 | Antibacterial Absorbable Surgical suture Polyglactin 910 1/2 circle reverse cutting 40mm OS Needle violet brided coated - 90cm, size- 1.with Triclosan coating.(Box of 12 Foils) | We would like request you to remove the "OS" from these products specification which is monopoly specification of other bidder. We request you to kindly remove the same from specification and accept the bid in general specification for Healthy competition. | | Antibacterial Absorbable Surgical suture Polyglactin 910 1/2 circle reverse cutting 40mm OS Needle violet brided coated - 90cm, size- 1.with Triclosan coating. | Not considered |
| | 2021101 | Oxidized Regenerated Cellulose based Topical Absorbable Hemostat, 1x2 inch (Box of 10 Foils) | We at this moment request you to accept the ISO/CE certificates in place of USFDA from Indian Manufacturer and allow the Indian bidders to participate in the tender to contribute towards MAKE IN INDIA campaign of PMO as per attached notification of PMO F.NO. Z.28018/67/2017-EPW Dtd:05.11.2019 | | Oxidized Regenerated Cellulose based Topical Absorbable Hemostat, Structured Non-Woven material, with Bactericidal property. Ease-of-use in both open and minimally invasive procedures. Approved by US FDA, sizes 1x2 inch | Not considered |
| | 2021102 | Oxidized Regenerated Cellulose based Topical Absorbable Hemostat, 2"x4" inch (Box of 10 Foils) | We at this moment request you to accept the ISO/CE certificates in place of USFDA from Indian Manufacturer and allow the Indian bidders to participate in the tender to contribute towards MAKE IN INDIA campaign of PMO as per attached notification of PMO F.NO. Z.28018/67/2017- | | Oxidized Regenerated Cellulose based Topical Absorbable Hemostat, Structured Non-Woven material, with Bactericidal property. Ease-of-use in both open and minimally invasive procedures. Approved by US FDA size 2"x4" inch | Not considered |

| Firm Name | Item Code | Item Name | Firm Remarks | Recommended Generic Item Name & Specification | As per Tender | Final Remarks |
|-----------|-----------|--------------------------------------|--|---|--|---|
| | | | EPW Dtd:05.11.2019 | | | |
| | 2021062 | Polyamide with cutting needle Size 1 | Please specify the complete specification of the tendered product. | | Size 1 | Deleted from the tender |
| | 2021063 | Polyamide with Loop Size 1 | Please specify the complete specification of the tendered product like needle length, sutures length, needle type etc | | Loop Size 1 | ½ circle round body needle 48mm ,suture length150 cms |
| | 10272 | Vicryl suture double Arm 7/0 | Request you to call the tender in generic name. Products called in brand name will restrict the participation of multiple bidders and encourage the monopoly bidding. Also please specify the complete specification of the tendered product like needle length, sutures length, needle type etc | | 7/0 Vicryl suture double Arm Port closure size should be 1 | 3/8 circle 7-0 double armed 6.5mm needle length, suture length 30cm |

| Firm Name | Item Code | Item Name | Firm Remarks | Recommended Generic Item Name & Specification | As per Tender | Final Remarks |
|---------------------------------|-----------|-----------|---|---|--|----------------|
| M/s Aegis Life Sciences Pvt Ltd | | | Firm requested for EMD and tender fees relaxation as the manufacturing unit is situated at Gujarat under category of MSME and NSIC. | | Only MSME of AP manufacturing units is exempted for EMD. | Not considered |