

**STATEMENT SHOWING REVISED PRILIMINARY TECHNICAL EVALUATION FOR T NO 163 (POC TEST KIT FOR SCREENING OF SICKLE CELL ANEMIA)**

S No	Bidder Name	Item Code	Item Name	Item Remarks	Status
1	<b>Alpine Biomedicals Private Limited</b>	<b>Technical Evaluation Remarks by DGDCA:</b> Applied item may be considered			
		<b>212611.01</b>	POC test kit for screening of sickle cell anemia		May be considered
2	<b>ANDHRA PRADESH MEDTECH ZONE LIMITED</b>	<b>Technical Evaluation Remarks by DGDCA:</b> The firm may be considered for item in S.No.1., after submission of <b>License copy of bidder</b> and <b>Non-Conviction Certificate of bidder.</b>			
		<b>212611.01</b>	POC test kit for screening of sickle cell anemia	Item may be considered after submission of Licence copy of the bidder and NCC of bidder	May not be considered
3	<b>MEDICLONE BIOTECH</b>	<b>Technical Evaluation Remarks by DGDCA:</b> The applied medical device catergerised under class 'C' which required license in Form MD-9 under medical Devices Rules 2017. Hence, the applied item may be considered after obtaining product approval license in Form MD-9 only			
		<b>212611.01</b>	POC test kit for screening of sickle cell anemia		May not be considered
4	<b>MERIL DIAGNOSTICS PVT. LTD</b>	<b>Technical Evaluation Remarks by DGDCA:</b> Applied item may be considered			
		<b>212611.01</b>	POC test kit for screening of sickle cell anemia		May be considered

S No	Bidder Name	Item Code	Item Name	Item Remarks	Status
5	ROYAL AGENCIES	<b>Technical Evaluation Remarks by DGDCA:</b> The applied medical device catergerised under class 'C' which required license in Form MD-9 under medical Devices Rules 2017. Hence, the applied item may be considered after obtaining product approval license in Form MD-9 only			
		212611.01	POC test kit for screening of sickle cell anemia		May not be considered
6	Sri Sai KrishnaMarketing Associates	<b>Technical Evaluation Remarks by DGDCA:</b> The applied Medical devices catergerised under class 'C' which required license in Form MD-9 under Medical devices Rules, 2017. Hence, the applied item may be considered after obtaining the product approval in Form MD-9 only			
		212611.01	POC test kit for screening of sickle cell anemia		May not be considered
7	TRANSASIA BIO-MEDICALS LTD	<b>Technical Evaluation Remarks by DGDCA:</b> The product in S.No.1 may be considered after submission of <b>NCC of bidder</b>			
		212611.01	POC test kit for screening of sickle cell anemia	Item may be considered after submission of NCC of bidder only	May not be considered
8	Unifranco Life Sciences Private Limited	<b>Technical Evaluation Remarks by DGDCA:</b> Applied item may be considered			
		212611.01	POC test kit for screening of sickle cell anemia		May be considered

S No	Bidder Name	Item Code	Item Name	Item Remarks	Status
9	SMPL LIFESCIENCES PRIVATE LIMITED	<b>Technical Evaluation Remarks by DGDCA:</b> The applied Medical devices catergerised under class 'C' as per Medical Devices Rules, 2017 which requires license under Form MD-9. Hence, the firm applied product may be considered after obtaining the product license in Form MD-9 only			
		212611.01	POC test kit for screening of sickle cell anemia		May not be considered
<p align="center"><b>Note: All the bidders are here by requested to verify the Provisional Technical Evaluation Report and submit their supporting documents <b>on or before 05.00PM on 12.08.2024</b> to the email id: <a href="mailto:tenders.apmsidc@gmail.com">tenders.apmsidc@gmail.com</a> with Subject : Reply to T No 163 (POC TEST KIT FOR SCREENING OF SICKLE CELL ANEMIA) TEC Report</b></p>					