

STATEMENT SHOWING PRILIMINARY TECHNICAL EVALUATION FOR T NO 163 (POC TEST KIT FOR SCREENING OF SICKLE CELL ANEMIA)

S No	Bidder Name	Item Code	Item Name	Item Remarks	Status
1	Alpine Biomedicals Private Limited	Technical Evaluation Remarks by DGDCA: Applied item may be considered			
		212611.01	POC test kit for screening of sickle cell anemia		May be considered
2	ANDHRA PRADESH MEDTECH ZONE LIMITED	Technical Evaluation Remarks by DGDCA: The firm may be considered for item in S.No.1., after submission of License copy of bidder and Non-Conviction Certificate of bidder.			
		212611.01	POC test kit for screening of sickle cell anemia	Item may be considered after submission of Licence copy of the bidder and NCC of bidder	May be considered
3	MEDICLONE BIOTECH	Technical Evaluation Remarks by DGDCA: The applied product may be considered after obtaining POCT recommendation of DG, ICMR for the applied item			
		212611.01	POC test kit for screening of sickle cell anemia	Item may be considered after submission of POCT recommendation of DG, ICMR/MOHFW	May be considered
4	MERIL DIAGNOSTICS PVT. LTD	Technical Evaluation Remarks by DGDCA: Applied item may be considered			
		212611.01	POC test kit for screening of sickle cell anemia		May be considered

S No	Bidder Name	Item Code	Item Name	Item Remarks	Status
5	ROYAL AGENCIES	Technical Evaluation Remarks by DGDCA: May be considered for product in S.No.1 after submission of valid/signed complete manufacturing license copy and POCT recommendation of DG, ICMR/MOHEW			
		212611.01	POC test kit for screening of sickle cell anemia	May be considered after submission of valid manufacturing license copy	May not be considered
6	Sri Sai KrishnaMarketing Associates	Technical Evaluation Remarks by DGDCA: (1) The firm may be considered after submission of POCT recommendation of DG, ICMR/MOHEW (2) Clarification in validation report of Rapid test kit (As in conclusion mentioned as showed 98.6% satisfactory match with HPLC gold standard results. Needs to improve band intensities of T1 and T2 position as they were found to very faint) (3) Submission of product approval as per Tender document i.e. POC Test kit for screening of SICKLE Cell ANEMIA (where as the bidder submitted approval for Sickle Chromatographic immunoarsy test)			
		212611.01	POC test kit for screening of sickle cell anemia	May be considered after submission of product approval as per tender docuement	May not be considered
7	TRANSASIA BIO-MEDICALS LTD	Technical Evaluation Remarks by DGDCA: The product in S.No.1 may be considered after submission of NCC of bidder			
		212611.01	POC test kit for screening of sickle cell anemia	Item may be considered after submission of NCC of bidder only	May be considered
8	Unifranco Life Sciences Private Limited	Technical Evaluation Remarks by DGDCA: The firm may be considered for product in S.No.1 after submission of POCT recommendation of DG, ICMR/MOHEW			

S No	Bidder Name	Item Code	Item Name	Item Remarks	Status
		212611.01	POC test kit for screening of sickle cell anemia	Item may be considered after submission of POCT recommendation of DG, ICMR/MOHFW	May be considered
9	SMPL LIFESCIENCES PRIVATE LIMITED	Technical Evaluation Remarks by DGDCA: The firm may be considered after submission of NCC of the bidder and POCT recommendation of DG, ICMR/MOHFW			
		212611.01	POC test kit for screening of sickle cell anemia	Item may be considered after submission of NCC of the bidder and POCT recommendation of DG, ICMR/MOHFW	May be considered
<p>Note: All the bidders are hereby requested to verify the Provisional Technical Evaluation Report and submit their supporting documents on or before 05.00PM on 01.08.2024 to the email id: tenders.apmsidc@gmail.com with Subject : Reply to T No 163 (POC TEST KIT FOR SCREENING OF SICKLE CELL ANEMIA) TEC Report</p>					