

Horrock's Apparatus

Should be suitable to find out the dose of bleaching powder (chlorine) required for disinfection of water.

Contents of apparatus:

1. 6 white cups with 200 ml capacity each.
2. 1 black cup
3. 2 metal spoons
4. 7 rods
5. 1 pipette
6. 2 droppers
7. Starch-iodine indicator solution

MUAC tapes

- Mid Upper Arm Circumference

Sound level meter

1. Multi-Function Environment Meter which has the functions of sound level meter
2. Sound level meter measuring range - 35 to 130dB, +/-3.5dB at 94db.
3. Light Meter measuring range - 20 to 20,000 lux, +/-5% reading.
4. Measuring Rate: 1.5 times per second.
5. Should have an LCD display for parameter display.
6. Should work on battery power source with charger.

Craft water testing kit

- BDI- E26 water and soil analysis kit(7 parameters)
- 3.5" digital LCD display, covering all 7 parameters, battery operated, with complete accessories
- ABS brief case

Public Address system (2 portable for field based activities and one each for RHTC &UHTC)

- Portable, Minimum area coverage up to 100 sq.m, Battery cum Mains Operated