

Final Technical Evaluation Report is herewith placed in website <https://msidc.ap.nic.in>. for all the bidders who have participated in Tender No.9.2H/APMSIDC/2023-24 Dt.06.02.2024, (Bids received on dt: 11.03.2024) for procurement and supply of Defibrillators to all DHs, AHs, CHCs under the control of DSH in Andhra Pradesh under 2 years Rate Contract.

Sd/-
Executive Director
APMSIDC

Date: 27.05.2024.

Mangalagiri, Guntur

As per Report submitted by technical committee after demonstration of equipment, details of evaluation report as follows.

Sl. No	Firm Name	Status of Final Evaluation Report
1.	M/s Pixel Imaging Systems Make & Model: Philips & Efficia DFM100	Recommended
2.	M/s Sai Srinivasa Agencies Make & Model: Nihan Kohden & TEC-5621	Recommended
3.	M/s Vishakha Healthcare India Pvt. Ltd Make & Model: Zoll & M2	Recommended
4.	M/s Instromedix India Pvt Ltd. Make & Model: Mediana Co. Ltd & D700	Self-test navigation is difficult and unable to do in emergency situation. Hence, not recommended.
5.	M/s Hyderabad Agricultural Co-operative Association Ltd. Make & Model: BPL & Relife-900	Subject to qualify, the committee informed to bidder to take declaration from OEM, on data transfer option (Data to PC through USB flash memory) at the time of demo. The bidder submitted declaration that, still manufacturing team working on the feasibility of data transfer through external port. Hence, not recommended.
6.	M/s Sri Vijaya Scientific Make & Model: CU Medical systems/CU HD1	Provisionally not qualified for demonstration
7.	M/s Allied Medical Ltd. Make & Model: Allied Medical Ltd, & Cardio safe-II	Provisionally not qualified for demonstration
8.	M/s SMPL Life Sciencies Pvt Ltd. Make & Model: Nasan & Life saver 108	Provisionally not qualified for demonstration