

Final Technical Evaluation Report is herewith placed in website <https://apmsidc.ap.nic.in> for all the bidders who have participated in Tender No.9.8A/APMSIDC/2023-24 Dt.06.02.2024, (Bids received on dt: 30.04.2024) for Procurement and Supply of Medical Equipment and Furniture to 2nd Phase 5 New Medical colleges/Hospitals in Andhra Pradesh – Tenders invited with 2years Rate Contract with reverse auction.

Sd/-
Executive Director
APMSIDC

Date: 15.07.2024.

Mangalagiri, Guntur

As per Report submitted by technical committee after demonstration of equipment, details of evaluation report as follows.

Table A: Final report on item wise Qualifications of Bids

Sl. No	Item Name	Firm name	Status of Final report
1	Group-1 (Ambu bag with face mask – Adult, Paediatric, Neonatal, Laryngoscope for Anterior & Oxygen head box)	M/s. Bhargav Enterprises	Not demonstrated
		M/s. Visakha Healthcare India Pvt. Ltd	Recommended
		M/s. CVR Medical systems	Not meeting specifications
		M/s. Ashwani Medical Pvt. Ltd	Recommended
2	Group-2 (Infantometer, Weighing Machine Digital – Child, Weighing Machine Digital – Infant & Weighing Machine Digital Neonate)	M/s. CVR Medical systems	Infantometer, weighing machine infant & neonatal not meeting specifications. Not Recommended
3	Vacuum Extractor and suction Machine	M/s. Bhargav Enterprises	Not demonstrated quoted model, given demonstration for another model. Not qualified
		M/s. Visakha Healthcare India Pvt. Ltd	Recommended
4	Baby incubator	M/s. Bhaskar Agency	Recommended
		M/s. Bhargav Enterprises	Recommended
5	HHHFNC	M/s. Bhaskar Agency	Not recommended
		M/s. Visakha Healthcare India Pvt. Ltd	Recommended
		M/s. The Hyderabad Agricultural Cooperative Association	Not meeting specifications
		M/s. J.S Industries	Quoted make & model not

Sl. No	Item Name	Firm name	Status of Final report
			shown
6	ETO steriliser	M/s. Bhargav Enterprises	Not qualified for demo
		M/s. The Hyderabad Agricultural Cooperative Association	Not demonstrated
7	Operation Theatre Ceiling light double Dome	M/s. Susruta Associates	Recommended
		M/s. Bhargav Enterprises	Not Qualified for demo
		M/s. Sai Srinivasa Agencies	Recommended
		M/s. SMPL Lifesciences Pvt. Ltd	Not Qualified for demo
		M/s. Hospedia Medicare Pvt. Ltd	Certificate of USFDA register / EUROPEAN CE (4 digit from a notified body) / bis approval from last 5 years or more not available (Not recommended)
8	Diathermy machine under water cutting	M/s. Colloq Marketing Pvt. Ltd	Not qualified for demo
		M/s. The Hyderabad Agricultural Cooperative Association	Not qualified for demo
		M/s. India Medtronic Pvt. Ltd	Recommended